

On this date 2/1/12, NPAS received instrument # 103006.
The instrument was sent from the Vermont Department of Health (VDH) so that NPAS could install updated software as well as inspect the instrument. At NPAS' request VDH provided copies of the Technical Support Inquiries (TSI) regarding the instrument. The calibration history of the instrument remained in the instrument so that NPAS could review its history.
NPAS examined the internal configuration of the instrument and did / did not find physical manipulation or tampering of the components located within the instrument itself.
NPAS did performance and accuracy checks on a certified 0.100 ethanol standard. NPAS found that the instrument passed failed the performance and accuracy checks.
NPAS reviewed the instrument as to its DataMaster DMT Inspection, QA and Conformance Testing Procedure and the instrument performed did not perform within its procedure.
NPAS examined the calibration records and did 7 did not observe anomalies.
NPAS testing indicates that there is no reason why this instrument would not have performed accurately and reliably in the field.
Print Name of Evaluator: Michael Barbely  Signature of Evaluator and Date: Mechael Barbely 2/3/12
Print Name of Reviewer: DAVID RADOMSKI
Signature of Reviewer and Date: Naud Radondhe 26/12

aborg 03/12/13 2m

Packing List

From: National Patent Analytical Systems Inc.

2090 Harrington Mem. Rd., Mansfield, Ohio 44903

Cust Purch Ord No. KIRK KIMBALL

Ship To:

Vermont Dept of Health

195 Colchester Avenue

Burlington

Attn:

Vt

05401

Ship Via FedEx

Comments Service for Instrument Serial Number 103006

Cust No. VT1678J Ord Date 2/1/12 Sales Order No. 45304

Packing List No.

Ship Date 2/8/12 69606

Bill To:

Vermont Dept of Health

25.00

P O Box 70

Burlington

Vt

Freight

05402

Inv Part No	Quantity Ordered	This Shipment	Back Ordered	Total Ship To Date	Description	C		In the second of the second	
	Ordered	Shipment	Ordered	10 Date		Customer Par	t No.	Drawing Part No. / Rev	
9DMT	1.0	1		1.0	DATAMASTER DMT SERVICE	RETURN			
S/N 103006	5					None Given	No	Draw # / NONE	
Serial Nos.	103006								1

Mores oblishis of